

# SIEVE RECERTIFICATION

## General Data

Serial number	1205-080
Sieve Aperture Size Purchased	220 $\mu\text{m}$
Original Manufacture Date	05/2016
Recertification Date	3/20/2014

## Measurement Results

## Tolerances

	Original	New		
Nominal Aperture Size	220 $\mu\text{m}$	220 $\mu\text{m}$	Max:	222 $\mu\text{m}$
			Min:	218 $\mu\text{m}$
Actual Aperture Size	220 $\mu\text{m}$	219.2 $\mu\text{m}$	Min:	222 $\mu\text{m}$
			Max:	218 $\mu\text{m}$

## Inspection Results

Frame Condition	Good
Epoxy Condition	Good
Wire Breaks	None
Support Grid/Bonding	Good
<b>FINAL RESULT</b>	<b>PASS</b>

Sieve medium satisfies the dimensional requirements of  
ASTM E11-09

The goods detailed were manufactured and inspected under a BS EN ISO 9001:2008

Measurement was carried out on optical inspection apparatus, in line with Annex A1 of  
ASTM E11-09

**PRECISION EFORMING** LLC

839 NYS Route 13

Cortland, NY 13045

TEL: 607.753.7730 FAX: 607.753.7680